

KAKEN

Kaken Test Center GENERAL INCORPORATED FOUNDATION

2-5-19, Edobori, Nishi-ku, Osaka, 550-0002/ Japan
Tel: +81-(0)6-6441-0315, Fax: +81-(0)6-6441-2420

Kaken Test Center is formerly named as "Japan Synthetic Textile Inspection Institute Foundation(JSTIIF)".

No.1/1

Certificate No. : OS-17-052767-4

Date : November 16, 2017

TEST CERTIFICATE

Requested : Dacian Technology Material Co., Ltd
Test Sample : dc face mask 1 sample
Test Item : Flammability
Received : November 14, 2017

This is to certify that the results of laboratory test applied on the sample are as follows:

1. Test Result

Test Item		Test Results		Test Method
		Time of Flame Spread (sec) or Burn Code		
		Original		
Flammability	1	DNI		16 CFR 1610
	2	DNI		
	3	DNI		
	4	DNI		
	5	DNI		
	AVG.			
Flammability Classification		Class 1		

Note) "DNI" stands for "Did Not Ignite."

2. Sample

KAKEN KAKEN KAKEN

KAKEN KAKEN KAKEN

Kaken Test Center General Incorporated Foundation
Osaka Laboratories

Period.

Inspector : T. Oida
T. Oida

Kaken Test Center
GENERAL INCORPORATED FOUNDATION
Lab Address: 2-5-19, Edobori, Nishi-ku, Osaka,
550-0002/ Japan
Tel: +81-(0)6-6441-0315
Fax: +81-(0)6-6441-2420

Any copying or replication of this report to or for any other person or entity, or use of our name or trademark, is permitted only with our prior written permission. This report sets forth our findings solely with respect to the test samples identified herein. The results set forth in this report are not indicative or representative of the quality or characteristics of the lot from which a test sample was taken or any similar or identical product unless specifically and expressly noted.